


Test 1 at RH20 with half of v2 board and v1 firmware

First Test run with one multiplexer

at ~20% RH on "old firmware" with single multiplexer



~20% RH

Device/sensor	t	RH
Green Day	23.6	19.6%
SHT20	23.85	22.06%
SHT20	23.86	21.92%
SHT20	23.80	21.55%
SHT21 (2)	23.80	21.99%
SHT21 (2)	23.90	22.59%
SHT21 (1)	23.89	32.81%

SHT21 (1)	23.83	30.93%
SHT25	24.02	22.01%
SHT30	23.93	23.54%
SHT30	23.90	23.10%
SHT31	23.86	24.50%
SHT31	23.82	24.49%
SHT31 (2)	23.87	26.41%
SHT31 (2)	23.81	22.12%
SHT35	23.85	22.17%
SHT35	23.85	22.41%
DHT12	23.40	30.40%
BME280	24.54	18.39%
BME280	24.28	18.71%
BME280	24.50	18.27%
BME680	24.70	21.41%
BME680	24.95	21.01%