




Temperature log [T_V8_009.zip](#)





Diff at ~65% RH

SHT20	SHT21-C	SHT21-G	SHT25	SHT30	SHT31	SHT31-2	SHT31-A	SHT35	SHTC1	SHTC3	SHT85
-0.19	0.32	-2.66	0.77	-0.01	0.62	-0.28	-0.10	0.28	1.10	-2.33	0.18
-0.21	0.83	-3.58		-0.12	0.71	1.07		0.14	1.29	-1.78	-0.26
0.26									1.41		0.09

BME280	BME280-G	DHT12	Si7021	Si7021-Y	Si7021-G	Si7021-A	HTU21d-Y	HTU21d-G	HTU21d-A
-5.37	-2.76	-4.65	-0.32	0.04	6.91	0.40	-3.40	-2.35	0.36
-4.89	-1.53	-10.25	-1.33	-1.74			-2.83	-2.86	
-1.74		-18.55		0.31			-2.59		
				0.85			-2.24		
							-2.84		

HDC1080-	HDC1080-	HDC2080-	AM2320-1	AM2320-2	AHT10	AHT15
8.89	11.86	0.93	11.45	6.15	2.44	-2.21
7.43	13.56	-0.10	12.55	13.95	19.77	-3.66
9.23	11.52	0.99	11.35	3.75	0.11	-0.75
	13.26	1.76		12.85	6.23	-2.01
	12.98	1.79		5.05		
		2.14				
		0.95				
		1.46				